



# Czech Metrology Institute

Notified Body No. 1383, Okružní 31  
638 00 Brno

## EC-TYPE EXAMINATION CERTIFICATE

Number: TCM 143/09 - 4664

### Addition 3

This addition replaces all previous versions of this certificate in full wording.

Issued by: **Český metrologický institut**  
**Okružní 31**  
**638 00 Brno**  
**Česká republika**

**Notified body no. 1383**

In accordance with: point 3 of annex 2 to Government Order No. 464/2005 Coll. (annex B of the Directive 2004/22/EC) from 19 October 2005 that lays down technical requirements on measuring instruments and implements in Czech Republic Directive 2004/22/EC of the European Parliament and of the Council.

Manufacturer: **ELGAS, s.r.o.**  
(Applicant) **Ohrazenice 211**  
**533 53 Pardubice**  
**Czech Republic**

In respect of: **volume conversion device**  
**MPE 0.5 %**  
**type: miniELCOR**  
**mechanical environment class: M2**  
**electromagnetic environment class: E2**  
**climatic environments limits: - 25°C...+ 70 °C**

Valid until: **14 April 2019**

Number of document: **0115-CS-A005-09**

Description: Essential characteristics, approved conditions and special conditions, if any, are described in this certificate. This certificate contains 13 pages.

Date of issue: 13 August 2013



.....  
**RNDr. Pavel Klenovský**

Notified Body No.1383

## 1. Characteristics of instrument

The volume conversion device miniELCOR (Fig.no.1 in annex) is an electronic gas volume conversion device which is designed to perform the continuous recalculation of volume at measurement conditions to the volume at base conditions. The volume at measurement conditions is obtained from pulse emitter or from Encoder interface of gas meter. The measurement of values of temperature and of pressure is performed in optional time periods. The volume conversion device miniELCOR is powered by batteries or by mains supply.

The conversion device can be produced:

- in the variant that consists of one temperature transducer and of one pressure transducer. The ratio of compressibility factors is calculated or the ratio of compressibility factors is constant.
- in the variant called T-convertor that consists of one temperature transducer and of no pressure transducer. The value of pressure is constant. The ratio of compressibility factors is calculated or the ratio of compressibility factors is constant.

The conversion device performs the recalculation of the volume at measurement conditions to the volume at base conditions

The base conditions can be:

$$p_b = 100.000 \text{ kPa or } 101.325 \text{ kPa or } 101.592 \text{ kPa or } 99.994664832 \text{ kPa (=14.503 psia) or}$$

$$101.325352987 \text{ kPa (14.696 psia) or } 101.559774734 \text{ kPa (=14.73 psia)}$$

$$t_b = 0^\circ\text{C or } 15^\circ\text{C (=59 F) or } 15.5555556^\circ\text{C (=60 F) or } 20^\circ\text{C or } 25^\circ\text{C}$$

These values  $p_p$  and  $t_b$  are fixed from manufacturer and the ones can be changed only after breaking of a seal (security mark).

In the software version 2.xx these equations are used for recalculations:

$$V_1 = \frac{N_1}{k_{p1}} \quad V_{b1} = V_1 * C_1 \quad C_1 = \frac{p_1}{p_b} * \frac{(t_b + 273.15)}{(t_1 + 273.15)} * \frac{1}{K_1}$$

In the software version 4.xx these equations can be used for recalculations:

$$V_m = \frac{N}{k_p} \quad V_b = V_m * C \text{ or } V_b = V_c * C \quad C = \frac{p}{p_b} * \frac{(t_b + 273.15)}{(t + 273.15)} * \frac{1}{K}$$

Symbol in the software version 2.xx	Symbol in the software version 4.xx	Represented quantity	Units
$C_1$	$C$	Conversion factor	-
$V_{b1}$	$V_b$	Volume at base conditions	$\text{m}^3$
$V_1$	$V_m$	Volume at measurement conditions	$\text{m}^3$
not applicable	$V_c$ (optionally)	Corrected volume according to the gas meter error	$\text{m}^3$
$V_s$	$V_s$ or $V_e$	Volume at measurement conditions during faulty condition(s)	$\text{m}^3$
$V_{bs}$	$V_{bs}$ or $V_{be}$	Volume at base conditions during faulty condition(s)	$\text{m}^3$
$N_1$	$N$	Number of pulses from gas meter	-
$k_{p1}$	$k_p$	Gas meter constant	$\text{imp}/\text{m}^3$
$p_1$	$p$	Absolute pressure at measurement conditions	kPa
$t_1$	$t$	Temperature at measurement conditions	$^\circ\text{C}$
$p_b$	$p_b$	Absolute pressure at base conditions	kPa
$t_b$	$t_b$	Temperature at base conditions	$^\circ\text{C}$





$K_f$	$K$ <i>(Z and <math>Z_b</math> can be also displayed.)</i>	The ratio of compressibility factors $K = \frac{Z}{Z_b}$ where $Z$ is the compressibility factor at measurement conditions $Z_b$ is the compressibility factor at base conditions	-
-------	---	--	---

In the software version 2.xx there the designations of quantities correspond to EN 12405-1/A1 with exception of  $V$ . (The symbol  $V_m$  is used for volume at measurement conditions in the harmonised standard.)

In the software version 4.xx there no index (1) is used. In this software version a corrected volume  $V_c$  according to the errors of gas meter in a defined flow rates can be optionally used if HF pulse emitter is used. Here the designations of quantities correspond to EN 12405-1/A1. Symbols  $V_e$  instead  $V_s$  and  $V_{be}$  instead  $V_{bs}$  can be optionally used in this software version 4.xx.

The units of values of pressures and of temperatures are optional and can be changed by user.

The conversion device miniELCOR counts the pulses from a gas meter and simultaneously the device measures the values of pressure and of temperature of the gas at measurement conditions by independent transducers. The transducers are inseparable from calculator so this device can be called gas volume conversion device type 1 (complete system - see 3.1.18.1. of EN 12405-1/A1).

The value  $K_f$  of the ratio of compressibility factors of natural gas can be computed by methods AGA NX-19 mod., AGA 8 DETAILED CHARACTERIZATION METHOD (designation AGA 8-92DC), SGERG-88, AGA 8 GROSS CHARACTERIZATION METHOD 1 (designation AGA 8-G1), AGA 8 GROSS CHARACTERIZATION METHOD 2 (designation AGA 8-G2), or the one can be constant ( $K_f=1$  or  $K_f \neq 1$ ). A user can change the method of calculation of the ratio of compressibility factors of natural gas but in case where the maximum pressure of the pressure transducer  $p_{max}$  is above 2000 kPa (20 bar) the temperature range can be reduced in accordance with the used method of compressibility calculation. The method AGA NX19 mod. is allowable only in the model with  $p_{max}=520$  kPa (5,2 bar). The change of the method of compressibility calculation is memorised in the setting archive. The actually used method of compressibility calculation is displayed on the LCD together with the corresponding temperature range.

The conversion device miniELCOR is also able to calculate the amount of energy but these functions were not assessed. Only if method AGA 8-92DC is used then the calorific value  $H_s$  is calculated according to EN ISO 6976 by the conversion device. A fixed value of  $H_s$  (changeable by user) is used for other methods. There is also possible to define reference conditions for calorific value  $H_s$  in the conversion device.

The conversion device miniELCOR consists of a calculator, of a temperature transducer (1 piece), of a pressure transducer (no pressure transducer or 1 piece). The mechanical construction allows connecting internal or external pressure transducers.

The electronics of the conversion device are placed on three boards. In the down part of the device body there is a board for inputs. This board contains also a main battery, a backup battery and terminal box for the connection of pressure sensor and of temperature sensor. Outputs are placed on this board, too. Optionally on this board the other board for connection of another external transducer of pressure or of temperature can be added. This optional transducer is not under scope of this type approval. In the upper part of device there is a board with processor. In the cover of this board there is an opening for the service switch which allows the access for parameters settings. (Fig.no.1 in annex).

There are two modifications of the conversion device miniELCOR:

- miniELCOR
- miniELCOR SCR1 - additional SCR interface is realized by "SCR board"

The equipment of the conversion device allows to measure the actual values of pressure  $p$  (or  $p_f$ ) at measurement conditions and of temperature  $t$  (or  $t_f$ ) at measurement conditions, to recalculate the volume at measurement conditions to the volume at base conditions, to summarize the increments to the counter of the volume at measurement conditions  $V_m$  (or  $V_f$ ) and to recalculate the increments to the counter of the volume at base conditions  $V_b$  (or  $V_{bf}$ ). In accordance with the parameter settings these and



other data are saved to archives. Furthermore the limits of values, defined alarms and the internal security are watched.

During the faulty conditions (a fault of the transducer, a fault of the device, a crossing of the limits) the volume at measurement conditions  $V_m$  (or  $V_l$ ) and the volume at base conditions  $V_b$  (or  $V_{bl}$ ) are collected to separate counters. The designation of the volumes at faulty conditions is  $V_s$  (or  $V_{sl}$  or  $V_e$ ) (at measurement conditions) and  $V_{bs}$  (or  $V_{bsl}$  or  $V_{be}$ ) (at base conditions). During the faulty conditions the device begins to count pulses using both counter of volume at measurement conditions  $V_m$  (or  $V_l$ ) and the counter for spare (substitute) volume at measurement conditions  $V_s$  (or  $V_{sl}$  or  $V_e$ ). During the faulty conditions the volume at base conditions  $V_b$  (or  $V_{bl}$ ) is not recorded and only the spare (substitute) volume at base conditions  $V_{bs}$  (or  $V_{bsl}$  or  $V_{be}$ ) is increased using spare (substitute) values for calculation. During faulty conditions no values are saved into the counter of volume at base conditions  $V_b$  (or  $V_{bl}$ ). When faulty conditions disappear then the saving to the counters of spare (substitute) volumes is stopped and the device starts to save volume at base condition  $V_b$  (or  $V_{bl}$ ) again.

The indication of values is performed using graphical LCD (128x64 points). The local control can be realised using six-button keyboard. Values to be displayed are selected by means of menu but the structure depends also on a configuration via the serial interface and software. To save energy of a power source the LCD is turned off after 20 seconds when the device is powered by battery. The LCD can be turned on by pressing of any button.

All data of actual values or of saved values or of parameters can be read and maintain by special software. The transition of data can be performed via interfaces (RS-232, RS-485 and IEC-1107 (HIE-01, HIE-03 a HIE-04)) using various protocols (ELGAS ver.2, MODBUS®). The communication with master system can be performed by various means – telephone modem, radio modem, GSM, GPRS. In this way various data from device can be read and it is possible to perform parameters setting.

Data influencing metrological characteristics of device are protected by metrological switch which is sealed. In the software version 2.xx there data which can be changed and which influence the calculation are protected by password and all changes are memorised in *setting archive*. In the software version 4.xx there parameters (composition) of natural gas are saved in the *gas changes archive* which is not erasable by user but cyclically overwrite able. It's possible to perform parameter changes via serial interface or via keyboard.

All changes of the configuration setting are saved. The allowed changes are mentioned down:

- values of volume at measurement conditions  $V_m$  (or  $V_l$ ) and of spare (substitute) volume at measurement conditions during faulty conditions  $V_s$  (or  $V_{sl}$  or  $V_e$ );
- the ratio of compressibility factors  $K$  (or  $K_l$ ) as a constant or the method of calculation of the ratio of compressibility  $K$  (or  $K_l$ )
- parameters of natural gas (In the software version 4.xx the changes are saved in the *gas changes archive*)
- pulse constant of gas meter  $k_p$  (or  $k_{pl}$ )
- units of quantities or of constants
- setting of archives, periods of measurements of quantities
- zeroing of archives with exception of the *setting archive*
- real time setting
- period of recalculation (Maximum period is 30 seconds.)
- spare (substitute) values of temperature and of pressure for measurement at faulty conditions
- the importance of service switch and the password setting

The device is powered by batteries. The device is able to work more then 5 years in the defined mode. After 90 % of battery lifetime an appropriate warning shall be shown. Data are saved when the define procedure of exchange of battery is used. If it is necessary to use the device in the mode with higher consumption of energy then the mains supply has to be used. The device also includes the standby battery which is capable to ensure the saving of the counters of volumes even during the interruption of the power supply. During the interruption of the power supply the circuits are powered by standby battery. That is why the spare (substitute) volumes are still saved and it is possible to continue to count input pulses from gas meter and to maintain the real time.

The absolute pressure transducer is used for pressure measurement. The main part is piezoresistive silicone sensor. The output signal from this sensor is changed in A/D transducer. The electronics perform the correction of non-linearity a of temperature dependence of pressure value.



The temperature is measured with two wires platinum resistance sensor Pt1000. The influence of the length and the properties of the sensor are taken into account during the calibration of the temperature sensor.

The volume at measurement condition is measured by counting the pulses from gas meter with defined pulse constant. The pulse input in the conversion device is able to evaluate pulses from LF emitters (reed contact, contact without potential, Wiegand) or from HF emitters (NAMUR, only with external power source).

The volume at measurement condition  $V_m$  (or  $V_f$ ) can be optionally obtained from gas meter using Encoder interface. In miniELCOR there is in this case a digital interface (based on NAMUR standard or SCR) which is capable to receive a totalizer number including checksum. An external power source has to be used in the case of using Encoder interface based on NAMUR standard. There is possible to connect Encoder interface based on SCR standard without using of an external power supply but in this case the battery lifetime is decreased.

The location of transducers of pressure and of temperature in the pipe(s) is recommended by procedure of the manufacturer. The conversion device is approved to be used in the hazardous area.

## 2. Main characteristics

Maximum permissible error of the conversion factor under reference conditions  $\pm 0.5 \%$

### Electromagnetic environments: E2

This class applies to instruments used in locations with electromagnetic disturbances corresponding to those likely to be found in other industrial buildings.

### Mechanical environments: M2

This class applies to instruments used in locations with significant or high levels of vibration and shock, e.g. transmitted from machines and passing vehicles in the vicinity or adjacent to heavy machines, conveyor belts, etc.

### Climatic environments: -25°C to 70°C

The device is designed for indoor areas with non-condensing humidity.

### Mechanical parameters:

- dimensions (width x height x depth)	(193 x 160 x 73) mm
- weight	maximally 1.2 kg
- material of casing	polycarbonate

Ingress protection: IP 66

### Explosion proof make:

- identification miniELCOR	 II 1G Ex ia IIC T4/T3
- area classification	ZONA 0, ZONA 1, ZONA 2
- identification miniELCOR SCR1	 II 2G Ex ib IIB T4/T3
- area classification	ZONA 1, ZONA 2
- certificate number	FTZÚ 08 ATEX 0324X

### Power supply:

- supplying battery type	lithium 3.6 V/17Ah
- supplying battery voltage	2.9 V – 3.6 V
- supplying battery life measuring	yes, warning after 90 % of its lifetime
- feeding from external power supply	from recommended source



- voltage of external JB power supply 4.5 V to 10 V
- standby battery type lithium 3.6 V/1Ah
- standby battery life 10 years in defined conditions

Software version: 2.xx

- 4.xx This software version is approved according to WELMEC 7.2 Guide:  
 Extension L: Long-term Storage of Measurement Data  
 Extension T: Transmission of Measurement Data via Communication Network  
 Extension D: Download of Legally Relevant Software

(The change of software version in digit place of integral is a subject of the supplement of this EC-type examination certificate. The changes of software version in the digit place of tenth and of hundredth do not change the metrology characteristics and they are allowed.)

The program calculates its own checksum and compares it with a saved value. Any error in the checksum will be detected and will result in an alarm.

SW remote download: A remote download of software version 4.xx is allowed. This remote download functionality was approved according to WELMEC 7.2 Guide – Extension D.

Period of measurement : 1 s to 30 s

Pressure measurement:

- number of inputs 0 or 1
- absolute pressure transducer silicon piezoresistive sensor
- type KP070
- measuring ranges (80 ÷ 250) kPa  
(80 ÷ 520) kPa  
(160 ÷ 520) kPa  
(200 ÷ 1000) kPa  
(300 ÷ 1000) kPa  
(400 ÷ 2000) kPa  
(700 ÷ 3500) kPa  
(1400 ÷ 7000) kPa  
(80 ÷ 1000) kPa  
(400 ÷ 7000) kPa
- accuracy of measurement < 0.25 % of a measured value
- maximal overload capacity 125 % of upper limit of measuring range
- make internal or external
- external sensor cable length maximum 5 m

Allowed setting of pressure constant (absolute pressure at measurement conditions) if the device is T-converter 80 kPa - 1000 kPa

Temperature measurement:

- number of inputs 1
- sensor resistive Pt1000, 2 wires
- type TR115
- measuring range -25°C to +60 °C



- measurement error	$\pm 0,2^{\circ}\text{C}$
- external sensor cable length	maximum 10 m
<u>Digital inputs:</u>	terminals: DI1-DI4
- number	4
- input options (SW configuration)	NF pulse input, HF pulse input, binary input
<u>NF pulse input:</u>	terminals: DI1-DI2
- maximal frequency	10 Hz
- input type	Reed contact or non-potential output, WIEGAND
- open circuit voltage	2.8 V – 3.6 V
<u>HF pulse input:</u>	terminals: DI1-DI2, with external power supply
- maximal frequency	5 kHz
- input type	NAMUR (DIN 19234)
- open circuit voltage	$U_{\text{PWR}} - 1 \text{ V}$
Encoder interface input: (only on an inputs board)	terminals: DI1 (HF+/-), with external power supply
- digital communication	based on Absolute-ENCODER/S1 protocol
- input type	NAMUR (DIN 19234)
Encoder interface input: (only on an inputs board)	terminals: A, B on SCR board, external power optional
- digital communication	based on EDIS and OBIS protocol
- input type	SCR+
<u>Interface for communication with master system:</u>	All interfaces share common communication channel.
Metallic interfaces	
- serial communication interface	RS-485 or RS-232
- communication protocol	optional, according to the firmware version
RS-232 line	terminals: TxD, RxD, GND1, CTS
RS-485 line	terminals: U1+, GND1, D1+, D1-
Interface IEC-1107	

The allowed ranges for methods of calculation of the ratio of compressibility factors:

Method	Pressure range	Temperature range
AGA NX-19 mod	(80 ÷ 250) kPa (80 ÷ 520) kPa (160 ÷ 520) kPa	(-25 ÷ +60)°C
AGA 8-92DC	(80 ÷ 250) kPa (80 ÷ 520) kPa (160 ÷ 520) kPa (200 ÷ 1000) kPa (300 ÷ 1000) kPa (400 ÷ 2000) kPa (700 ÷ 3500) kPa (1400 ÷ 7000) kPa (80 ÷ 1000) kPa (400 ÷ 7000) kPa	(-25 ÷ +60)°C
SGERG-88 AGA 8-G1 AGA 8-G2	(80 ÷ 250) kPa (80 ÷ 520) kPa (160 ÷ 520) kPa (200 ÷ 1000) kPa (300 ÷ 1000) kPa (400 ÷ 2000) kPa (80 ÷ 1000) kPa	(-25 ÷ +60)°C
SGERG-88 AGA 8-G1 AGA 8-G2	(700 ÷ 3500) kPa (1400 ÷ 7000) kPa (400 ÷ 7000) kPa	(-10 ÷ +60)°C

Note: If a measured value of temperature is out of limits of the used method of calculation of the ratio of compressibility factors then the device starts to count the volumes into spare (substitute) volumes. In this case the spare (substitute) value of the ratio of compressibility factors is used but the actual value of temperature and the actual value of pressure are used if they are in the ranges of used transducers.

### 3. Test

Tests and conformity assessment according to the module B for addition 3 were performed in the laboratory of Czech Metrology Institute (department of gas flow) and in the laboratory of manufacturer. Serial number of tested sample was 1309001003.

Tests and assessments were performed according to Annex 2 of the harmonised standard EN 12405-1:2005/A1:2006 *Gas meters – Conversion devices- Part 1: Volume conversion*. All used standard meters were traceable to national standards and were regularly recalibrated.

The conversion device of type miniELCOR complied with all specified requirements and the one is able to fulfil the determined purpose which was designed for.

The results of tests and of assessment are summarised in the *Test Report No. 5012-PT-A0006-13*.

### 4. Markings

Each conversion device shall be marked with the following information on the device (Fig. no.2 in annex) :

- type: miniELCOR
- serial number/ year of manufacture
- name of manufacturer
- MPE at reference conditions
- ingress protection (IP code)
- number of the EC-type examination certificate: **TCM 143/09-4664**
- range of the pressure transducer
- maximal range of the temperature transducer





- identification mark for using in hazardous area  $\text{Ex}$  II 1G Ex ia IIC T4/T3 or  $\text{Ex}$  II 2G Ex ib IIB T4/T3
- valid certificate number for explosion (hazardous) area FTZÚ 08 ATEX 0324X
- extreme temperatures of the environmental class  $T_{amb}$
- 'CE' marking and supplementary metrology marking 'M' and the last two digits of the year of its affixing. The identification number of the notified body shall follow the 'CE' marking and supplementary metrology marking.

On the display (LCD) there the following information shall be indicated:

- base conditions  $t_b$  and  $p_b$
- actual method of calculation of the ratio of compressibility factors
- actual range of the temperature transducer
- parameters (composition) of natural gas if the ratio of compressibility factors  $K_f$  is not constant (or the value of  $K$ , if  $K$  is constant)
- value(s) of pulse constant(s) of gas meter  $k_p$  (imp/m<sup>3</sup>)

The language used on labels can be the language of the destination where the device is produced for.

The conversion device that corresponds to this EC-type examination certificate and to other requirements concerning the assessment according to the module F or D is sealed in the way mentioned in Fig. no. 3 and in Fig. no. 4 in annex. A metrological switch is switched off before sealing. The adhesive stickers are used for sealing. Place of sealing:

- main device label ... 1 sticker
- cover on PCB in the top of the conversion device ... 2 stickers
- label covering the metrology switch ... 1 sticker
- cover on PCB in the bottom of the conversion device ... 1 sticker
- cover on SCR board in the bottom of the conversion device  
(Only if an SCR board is used.) ... 1 sticker

#### Annex:

- Figure no.1 Design of the conversion device miniELCOR
- Figure no.2 Czech and English variants of labels on the front (top) panel of miniELCOR (Other language variations are allowed.)
- Figure no.3 Places of seals (security marks) in form of stickers on the conversion device of type miniELCOR (not for variant SCR1)
- Figure no.4 Places of seals (security marks) in form of stickers on the conversion device of type miniELCOR SCR1


















Figure no.1 Design of the conversion device miniELCOR



 <b>miniELCOR</b> 1009002061 r.v.: 2010 <b>Přepočítávač množství plynu</b> Tlak: (80 ÷ 520) kPa Teplota: (-25 ÷ 60) °C (-25 ≤ T <sub>amb</sub> ≤ 70) °C MPE: ± 0.5 % ELGAS, s.r.o. Pardubice Czech Republic	Certifikát ES přezkoušení typu: TCM 143/09 - 4664 FTZÚ 08 ATEX 0324X IP 66 (Ex) II 1G Ex ia IIC T4/T3 T4: (-25 ≤ T <sub>amb</sub> ≤ 40) °C T3: (-25 ≤ T <sub>amb</sub> ≤ 70) °C    1383 1026
<b>Pozor! Nebezpečí elektrostatického nabití - netřít!</b>	

 <b>miniELCOR</b> 1009000758 Date: 2010 <b>Volume Conversion Device</b> Pressure: (80 ÷ 520) kPa Temperature: (-25 ÷ 60) °C (-25 ≤ T <sub>amb</sub> ≤ 70) °C MPE: ± 0.5 % ELGAS, s.r.o. Pardubice Czech Republic	EC-type examination certificate: TCM 143/09 - 4664 FTZÚ 08 ATEX 0324X IP 66 (Ex) II 1G Ex ia IIC T4/T3 T4: (-25 ≤ T <sub>amb</sub> ≤ 40) °C T3: (-25 ≤ T <sub>amb</sub> ≤ 70) °C    1383 1026
<b>ATTENTION! Electrostatic hazard. Do not rub.</b>	

 <b>miniELCOR SCR1</b> 1009000758 Date: 2010 <b>Volume Conversion Device</b> Pressure: (80 ÷ 520) kPa Temperature: (-25 ÷ 60) °C (-25 ≤ T <sub>amb</sub> ≤ 70) °C MPE: ± 0.5 % ELGAS, s.r.o. Pardubice Czech Republic	EC-type examination certificate: TCM 143/09 - 4664 FTZÚ 08 ATEX 0324X IP 66 (Ex) II 2G Ex ib IIB T4/T3 T4: (-25 ≤ T <sub>amb</sub> ≤ 40) °C T3: (-25 ≤ T <sub>amb</sub> ≤ 70) °C    1383 1026
<b>ATTENTION! Electrostatic hazard. Do not rub.</b>	





 <b>miniELCOR SCR1</b> 1009002061 r.v.: 2010 <b>Přepočítávač množství plynu</b> Tlak: (80 ÷ 520) kPa Teplota: (-25 ÷ 60) °C (-25 ≤ T <sub>amb</sub> ≤ 70) °C MPE: ± 0.5 % ELGAS, s.r.o. Pardubice Czech Republic	Certifikát ES přezkoušení typu: TCM 143/09 - 4664 FTZÚ 08 ATEX 0324X IP 66 (Ex) II 2G Ex ib IIB T4/T3 T4: (-25 ≤ T <sub>amb</sub> ≤ 40) °C T3: (-25 ≤ T <sub>amb</sub> ≤ 70) °C    1383 1026
<b>Pozor! Nebezpečí elektrostatického nabití - netřít!</b>	

Figure no.2 Czech and English variants of labels on the front (top) panel of miniELCOR (Other language variations are allowed.)








Figure no.3 Places of seals (security marks) in form of stickers on the conversion device of type miniELCOR (not for variant SCR1)




Figure no.4 Places of seals (security marks) in form of stickers on the conversion device of type miniELCOR SCR1